

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00691	Calib. Date: 07/28/2017
Certification File No.:	00692	Calib. No.: 00024
Linearity File No.:	00693	Cert. Date: 07/28/2017
Solution File No.:	00715	Cert. No.: 00020
Sequential File No.:	00715	Lin. Date: 07/28/2017
		Lin. No.: 00020
		Soln. Date: 12/21/2017
		Soln. No.: 00148
		File Date: 12/21/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0052
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 0107

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:12S		
Control 1 EC	0.101%	17:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14S		
Control 2 EC	0.100%	17:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:15S		
Control 3 EC	0.100%	17:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____



Badge No.: 161

Date: 12/21/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0073
Location:	BRIELLE POLICE DEPT.		
Calibration File No.:	00691	Calib. Date:	07/28/2017
Certification File No.:	00692	Cert. Date:	07/28/2017
Linearity File No.:	00693	Lin. Date:	07/28/2017
Solution File No.:	00711	Soln. Date:	11/28/2017
Sequential File No.:	00711	File Date:	11/28/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0052
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	0517

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:17S		
Control 1 EC	0.102%	20:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:18S		
Control 2 EC	0.101%	20:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:20S		
Control 3 EC	0.101%	20:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature:  #161

Badge No.: 161

Date: 11/28/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00691	Calib. Date: 07/28/2017
Certification File No.:	00692	Calib. No.: 00024
Linearity File No.:	00693	Cert. Date: 07/28/2017
Solution File No.:	00710	Cert. No.: 00020
Sequential File No.:	00710	Lin. Date: 07/28/2017
		Lin. No.: 00020
		Soln. Date: 11/03/2017
		Soln. No.: 00146
		File Date: 11/03/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0052
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 0629

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:07S		
Control 1 EC	0.100%	09:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:08S		
Control 2 EC	0.100%	09:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:09S		
Control 3 EC	0.100%	09:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____



Badge No.: 161

Date: 11/03/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0073
Location:	BRIELLE POLICE DEPT.		
Calibration File No.:	00691	Calib. Date:	07/28/2017
Certification File No.:	00692	Cert. No.:	00024
Linearity File No.:	00693	Cert. Date:	07/28/2017
Solution File No.:	00704	Lin. Date:	07/28/2017
Sequential File No.:	00704	Lin. No.:	00020
		Soln. Date:	09/25/2017
		Soln. No.:	00145
		File Date:	09/25/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0052
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	0621

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:08D		
Control 1 EC	0.100%	06:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:09D		
Control 2 EC	0.099%	06:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10D		
Control 3 EC	0.099%	06:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: 

Badge No.: 161

Date: 09/25/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00691 Calib. Date: 07/28/2017 Calib. No.: 00024
Certification File No.: 00692 Cert. Date: 07/28/2017 Cert. No.: 00020
Linearity File No.: 00693 Lin. Date: 07/28/2017 Lin. No.: 00020
Solution File No.: 00699 Soln. Date: 08/25/2017 Soln. No.: 00144
Sequential File No.: 00699 File Date: 08/25/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0052
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 0108

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:24D		
Control 1 EC	0.099%	11:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		
Control 2 EC	0.098%	11:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		
Control 3 EC	0.099%	11:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: PH. [Signature] #161

Badge No.: 161

Date: 08/25/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00660 **Calib. Date:** 02/09/2017 **Calib. No.:** 00023
Certification File No.: 00661 **Cert. Date:** 02/09/2017 **Cert. No.:** 00019
Linearity File No.: 00662 **Lin. Date:** 02/09/2017 **Lin. No.:** 00019
Solution File No.: 00690 **Soln. Date:** 07/06/2017 **Soln. No.:** 00142
Sequential File No.: 00690 **File Date:** 07/06/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXA S3-0051
Control Solution %: 0.100% **Expires:** 01/30/2019
Solution Control Lot: 17050 **Bottle No.:** 0487

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:35D		
Control 1 EC	0.102%	08:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:36D		
Control 2 EC	0.101%	08:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37D		
Control 3 EC	0.100%	08:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: 

Badge No.: 161

Date: 07/06/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00660 Calib. Date: 02/09/2017 Calib. No.: 00023
Certification File No.: 00661 Cert. Date: 02/09/2017 Cert. No.: 00019
Linearity File No.: 00662 Lin. Date: 02/09/2017 Lin. No.: 00019
Solution File No.: 00684 Soln. Date: 06/05/2017 Soln. No.: 00141
Sequential File No.: 00684 File Date: 06/05/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0051
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 0148

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:00D		
Control 1 EC	0.102%	05:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:01D		
Control 2 EC	0.101%	05:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	05:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:02D		
Control 3 EC	0.101%	05:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	05:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: 

Badge No.: 161

Date: 06/05/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00660	Calib. Date: 02/09/2017
Certification File No.:	00661	Calib. No.: 00023
Linearity File No.:	00662	Cert. Date: 02/09/2017
Solution File No.:	00679	Cert. No.: 00019
Sequential File No.:	00679	Lin. Date: 02/09/2017
		Lin. No.: 00019
		Soln. Date: 05/02/2017
		Soln. No.: 00140
		File Date: 05/02/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 0157

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:13D		
Control 1 EC	0.100%	14:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14D		
Control 2 EC	0.100%	14:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15D		
Control 3 EC	0.100%	14:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 05/02/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00660	Calib. Date: 02/09/2017
Certification File No.:	00661	Calib. No.: 00023
Linearity File No.:	00662	Cert. Date: 02/09/2017
Solution File No.:	00676	Cert. No.: 00019
Sequential File No.:	00676	Lin. Date: 02/09/2017
		Lin. No.: 00019
		Soln. Date: 04/05/2017
		Soln. No.: 00139
		File Date: 04/05/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 0152

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:12D		
Control 1 EC	0.101%	04:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	04:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:13D		
Control 2 EC	0.101%	04:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	04:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:14D		
Control 3 EC	0.101%	04:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	04:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 04/05/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0073
Location:	BRIELLE POLICE DEPT.	Calib. Date:	02/09/2017
Calibration File No.:	00660	Calib. No.:	00023
Certification File No.:	00661	Cert. Date:	02/09/2017
Linearity File No.:	00662	Lin. Date:	02/09/2017
Solution File No.:	00670	Soln. Date:	03/07/2017
Sequential File No.:	00670	File Date:	03/07/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0051
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0261

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:44S		
Control 1 EC	0.103%	13:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	13:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 2 EC	0.102%	13:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	13:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47S		
Control 3 EC	0.102%	13:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	13:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSEN

First Name: LT.

MI: -

Signature: _____

Badge No.: 154

Date: 03/07/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00660 **Calib. Date:** 02/09/2017 **Calib. No.:** 00023
Certification File No.: 00661 **Cert. Date:** 02/09/2017 **Cert. No.:** 00019
Linearity File No.: 00662 **Lin. Date:** 02/09/2017 **Lin. No.:** 00019
Solution File No.: 00663 **Soln. Date:** 02/09/2017 **Soln. No.:** 00137
Sequential File No.: 00663 **File Date:** 02/09/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXA S3-0051
Control Solution %: 0.100% **Expires:** 05/17/2018
Solution Control Lot: 201605D **Bottle No.:** 0263

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:55S		
Control 1 EC	0.104%	09:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		
Control 2 EC	0.103%	09:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		
Control 3 EC	0.103%	09:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDW3P2-227 (DK)

Changed By:

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature:  #7040

Badge No.: 7040

Date: 02/09/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00645	Calib. Date: 08/26/2016
Certification File No.:	00646	Calib. No.: 00022
Linearity File No.:	00647	Cert. Date: 08/26/2016
Solution File No.:	00657	Cert. No.: 00018
Sequential File No.:	00657	Lin. Date: 08/26/2016
		Lin. No.: 00018
		Soln. Date: 01/16/2017
		Soln. No.: 00136
		File Date: 01/16/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	201605D	Expires: 05/17/2018
		Bottle No.: 0265

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:40S		
Control 1 EC	0.102%	12:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41S		
Control 2 EC	0.101%	12:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43S		
Control 3 EC	0.101%	12:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSEN

First Name: LT.

MI: -

Signature: _____

Badge No.: 154

Date: 01/16/2017