

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00760 Calib. Date: 08/13/2018 Calib. No.: 00029
Certification File No.: 00761 Cert. Date: 08/13/2018 Cert. No.: 00024
Linearity File No.: 00762 Lin. Date: 08/13/2018 Lin. No.: 00023
Solution File No.: 00776 Soln. Date: 12/18/2018 Soln. No.: 00163
Sequential File No.: 00776 File Date: 12/18/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0052
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 1195

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:11S		
Control 1 EC	0.100%	23:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:12S		
Control 2 EC	0.099%	23:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:13S		
Control 3 EC	0.099%	23:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____



Badge No.: 161

Date: 12/18/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00760	Calib. Date: 08/13/2018
Certification File No.:	00761	Calib. No.: 00029
Linearity File No.:	00762	Cert. Date: 08/13/2018
Solution File No.:	00775	Cert. No.: 00024
Sequential File No.:	00775	Lin. Date: 08/13/2018
		Lin. No.: 00023
		Soln. Date: 11/24/2018
		Soln. No.: 00162
		File Date: 11/24/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0052
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0902

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:35S		
Control 1 EC	0.099%	09:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36S		
Control 2 EC	0.098%	09:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37S		
Control 3 EC	0.098%	09:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 11/24/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0073
Location:	BRIELLE POLICE DEPT.		
Calibration File No.:	00760	Calib. Date:	08/13/2018
Certification File No.:	00761	Cert. Date:	08/13/2018
Linearity File No.:	00762	Lin. Date:	08/13/2018
Solution File No.:	00772	Soln. Date:	10/29/2018
Sequential File No.:	00772	File Date:	10/29/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0052
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	0793

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:54S		
Control 1 EC	0.101%	03:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:55S		
Control 2 EC	0.099%	03:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:56S		
Control 3 EC	0.100%	03:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 10/29/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00760	Calib. Date: 08/13/2018
Certification File No.:	00761	Calib. No.: 00029
Linearity File No.:	00762	Cert. Date: 08/13/2018
Solution File No.:	00768	Cert. No.: 00024
Sequential File No.:	00768	Lin. Date: 08/13/2018
		Lin. No.: 00023
		Soln. Date: 10/04/2018
		Soln. No.: 00160
		File Date: 10/04/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0052
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0724

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:03D		
Control 1 EC	0.101%	08:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04D		
Control 2 EC	0.100%	08:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:05D		
Control 3 EC	0.100%	08:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____



Badge No.: 161

Date: 10/04/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00760	Calib. Date: 08/13/2018
Certification File No.:	00761	Calib. No.: 00029
Linearity File No.:	00762	Cert. Date: 08/13/2018
Solution File No.:	00766	Cert. No.: 00024
Sequential File No.:	00766	Lin. Date: 08/13/2018
		Lin. No.: 00023
		Soln. Date: 09/07/2018
		Soln. No.: 00159
		File Date: 09/07/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0052
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1022

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:09D		
Control 1 EC	0.100%	06:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10D		
Control 2 EC	0.099%	06:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11D		
Control 3 EC	0.099%	06:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____



Badge No.: 161

Date: 09/07/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00760	Calib. Date: 08/13/2018
Certification File No.:	00761	Calib. No.: 00029
Linearity File No.:	00762	Cert. Date: 08/13/2018
Solution File No.:	00763	Cert. No.: 00024
Sequential File No.:	00763	Lin. Date: 08/13/2018
		Lin. No.: 00023
		Soln. Date: 08/13/2018
		Soln. No.: 00158
		File Date: 08/13/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0052
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0785

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:20D		
Control 1 EC	0.100%	14:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21D		
Control 2 EC	0.100%	14:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:22D		
Control 3 EC	0.100%	14:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Temperature Probe Serial Number: DDWJ P2-228A

Changed By:

Last Name: PAVLOSKY

First Name: ALLISON

MI: M.

Signature: 

Badge No.: 7330

Date: 08/13/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 01/19/2018
Certification File No.:	00721	Calib. No.: 00025
Linearity File No.:	00722	Cert. Date: 01/19/2018
Solution File No.:	00746	Cert. No.: 00021
Sequential File No.:	00746	Lin. Date: 01/19/2018
		Lin. No.: 00021
		Soln. Date: 06/10/2018
		Soln. No.: 00156
		File Date: 06/10/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1031

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:15D		
Control 1 EC	0.100%	16:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:16D		
Control 2 EC	0.099%	16:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17D		
Control 3 EC	0.100%	16:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 06/10/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 01/19/2018
Certification File No.:	00721	Calib. No.: 00025
Linearity File No.:	00722	Cert. Date: 01/19/2018
Solution File No.:	00741	Cert. No.: 00021
Sequential File No.:	00741	Lin. Date: 01/19/2018
		Lin. No.: 00021
		Soln. Date: 05/18/2018
		Soln. No.: 00155
		File Date: 05/18/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0853

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:31D		
Control 1 EC	0.099%	04:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:32D		
Control 2 EC	0.099%	04:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:34D		
Control 3 EC	0.099%	04:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	04:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOKFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 05/18/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0073
Location:	BRIELLE POLICE DEPT.		
Calibration File No.:	00720	Calib. Date:	01/19/2018
Certification File No.:	00721	Cert. Date:	01/19/2018
Linearity File No.:	00722	Lin. Date:	01/19/2018
Solution File No.:	00740	Soln. Date:	04/24/2018
Sequential File No.:	00740	File Date:	04/24/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0051
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	1033

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:48D		
Control 1 EC	0.099%	19:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:49D		
Control 2 EC	0.097%	19:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:50D		
Control 3 EC	0.097%	19:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: P.H. [Signature] #161

Badge No.: 161

Date: 04/24/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 01/19/2018
Certification File No.:	00721	Calib. No.: 00025
Linearity File No.:	00722	Cert. Date: 01/19/2018
Solution File No.:	00732	Cert. No.: 00021
Sequential File No.:	00732	Lin. Date: 01/19/2018
		Lin. No.: 00021
		Soln. Date: 03/31/2018
		Soln. No.: 00153
		File Date: 03/31/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0832

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:49S		
Control 1 EC	0.100%	15:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51S		
Control 2 EC	0.098%	15:51S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:52S		
Control 3 EC	0.099%	15:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 03/31/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 01/19/2018
Certification File No.:	00721	Calib. No.: 00025
Linearity File No.:	00722	Cert. Date: 01/19/2018
Solution File No.:	00730	Cert. No.: 00021
Sequential File No.:	00730	Lin. Date: 01/19/2018
		Lin. No.: 00021
		Soln. Date: 03/08/2018
		Soln. No.: 00152
		File Date: 03/08/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1286

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:25S		
Control 1 EC	0.098%	05:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:26S		
Control 2 EC	0.098%	05:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28S		
Control 3 EC	0.098%	05:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Badge No.: 161

Date: 03/08/2018

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 01/19/2018
Certification File No.:	00721	Calib. No.: 00025
Linearity File No.:	00722	Cert. Date: 01/19/2018
Solution File No.:	00728	Cert. No.: 00021
Sequential File No.:	00728	Lin. Date: 01/19/2018
		Lin. No.: 00021
		Soln. Date: 02/10/2018
		Soln. No.: 00151
		File Date: 02/10/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1277

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:48S		
Control 1 EC	0.098%	14:49S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:50S		
Control 2 EC	0.097%	14:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51S		
Control 3 EC	0.098%	14:52S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STAHL

First Name: BRIAN

MI: C

Signature: _____

[Handwritten Signature] #165

Badge No.: 165

Date: 02/10/2018

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 01/19/2018
Certification File No.:	00721	Cert. No.: 00025
Linearity File No.:	00722	Lin. No.: 00021
Solution File No.:	00723	Soln. No.: 00150
Sequential File No.:	00723	File Date: 01/19/2018

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 0766

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:36S		
Control 1 EC	0.103%	14:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38S		
Control 2 EC	0.101%	14:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39S		
Control 3 EC	0.101%	14:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2-227 (DK)

Changed By:

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: T. I. D. KLIMIK #7040

Badge No.: 7040

Date: 01/19/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00691 Calib. Date: 07/28/2017 Calib. No.: 00024
Certification File No.: 00692 Cert. Date: 07/28/2017 Cert. No.: 00020
Linearity File No.: 00693 Lin. Date: 07/28/2017 Lin. No.: 00020
Solution File No.: 00719 Soln. Date: 01/16/2018 Soln. No.: 00149
Sequential File No.: 00719 File Date: 01/16/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0052
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 0500

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:35S		
Control 1 EC	0.102%	03:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:36S		
Control 2 EC	0.100%	03:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:38S		
Control 3 EC	0.101%	03:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature:  #161

Badge No.: 161

Date: 01/16/2018