

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Brielle Police Department 601 Union Lane Brielle, NJ 08730</b>		2. Contact: <b>Ron Sofield</b>	4. Date: <b>9/1/20</b>
		3. Phone Number: <b>732-528-5050</b>	5. County: <b>Monmouth</b>
6. Alcotest Instrument Serial Number: <b>NA</b>	7. Simulator Component Serial Number: <b>DDXA S3-0052</b>	8. Temperature Probe Component Serial Number: <b>DDWJP2-228</b>	

9. Reason for Service:  
**I responded to Brielle Police Department to complete the semi-annual recalibration of the Alcotest 7110. The simulator (Serial #DDXA S3-0052) was not heating. The light would come on for a brief time after powering on and then go out and remain out. The solution would not warm up.**

10. Comments:  
**The simulator and probe were placed out of service and returned to Draeger Inc for evaluation.**

Last Known Sequential File #: NA

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045  
Name & Badge Number (Print)

Tpr I Lutz 7045 9-1-20  
Signature Date