

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00893 Calib. Date: 02/16/2021 Calib. No.: 00035
Certification File No.: 00894 Cert. Date: 02/16/2021 Cert. No.: 00030
Linearity File No.: 00895 Lin. Date: 02/16/2021 Lin. No.: 00029
Solution File No.: 00898 Soln. Date: 02/22/2021 Soln. No.: 00206
Sequential File No.: 00898 File Date: 02/22/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0051
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 0060

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:42S		
Control 1 EC	0.099%	10:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44S		
Control 2 EC	0.096%	10:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46S		
Control 3 EC	0.097%	10:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: Lt. Stephen Boyd #156

Badge No.: 156

Date: 02/22/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00893 Calib. Date: 02/16/2021 Calib. No.: 00035
Certification File No.: 00894 Cert. Date: 02/16/2021 Cert. No.: 00030
Linearity File No.: 00895 Lin. Date: 02/16/2021 Lin. No.: 00029
Solution File No.: 00900 Soln. Date: 03/12/2021 Soln. No.: 00207
Sequential File No.: 00900 File Date: 03/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0051
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 1174

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:54S		
Control 1 EC	0.099%	08:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55S		
Control 2 EC	0.097%	08:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57S		
Control 3 EC	0.098%	08:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: Lt. 

Badge No.: 156

Date: 03/12/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00893	Calib. Date: 02/16/2021
Certification File No.:	00894	Calib. No.: 00035
Linearity File No.:	00895	Cert. Date: 02/16/2021
Solution File No.:	00903	Cert. No.: 00030
Sequential File No.:	00903	Lin. Date: 02/16/2021
		Lin. No.: 00029
		Soln. Date: 03/15/2021
		Soln. No.: 00209
		File Date: 03/15/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1106

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:39S		
Control 1 EC	0.100%	10:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41S		
Control 2 EC	0.097%	10:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43S		
Control 3 EC	0.098%	10:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: _____



Badge No.: 156

Date: 03/15/2021

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00937 Calib. Date: 10/25/2021 Calib. No.: 00038
Certification File No.: 00938 Cert. Date: 10/25/2021 Cert. No.: 00032
Linearity File No.: 00939 Lin. Date: 10/25/2021 Lin. No.: 00031
Solution File No.: 00945 Soln. Date: 11/19/2021 Soln. No.: 00221
Sequential File No.: 00945 File Date: 11/19/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0051
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0724

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:35S		
Control 1 EC	0.102%	16:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:36S		
Control 2 EC	0.100%	16:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	16:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:37S		
Control 3 EC	0.101%	16:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	16:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:39S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: 

Badge No.: 156

Date: 11/19/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00906 Calib. Date: 04/26/2021 Calib. No.: 00037
Certification File No.: 00907 Cert. Date: 04/26/2021 Cert. No.: 00031
Linearity File No.: 00908 Lin. Date: 04/26/2021 Lin. No.: 00030
Solution File No.: 00915 Soln. Date: 06/23/2021 Soln. No.: 00214
Sequential File No.: 00915 File Date: 06/23/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0052
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0723

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:53D		
Control 1 EC	0.100%	14:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		
Control 2 EC	0.099%	14:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56D		
Control 3 EC	0.099%	14:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEVE

MI: M

Signature: Lt.  #156

Badge No.: 156

Date: 06/23/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00906 Calib. Date: 04/26/2021 Calib. No.: 00037
Certification File No.: 00907 Cert. Date: 04/26/2021 Cert. No.: 00031
Linearity File No.: 00908 Lin. Date: 04/26/2021 Lin. No.: 00030
Solution File No.: 00926 Soln. Date: 08/13/2021 Soln. No.: 00217
Sequential File No.: 00926 File Date: 08/13/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0052
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0726

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:35D		
Control 1 EC	0.099%	09:35D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36D		
Control 2 EC	0.099%	09:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37D		
Control 3 EC	0.099%	09:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: Lt. Stephen Boyd #156

Badge No.: 156

Date: 08/13/2021

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00937	Calib. Date: 10/25/2021
Certification File No.:	00938	Calib. No.: 00038
Linearity File No.:	00939	Cert. Date: 10/25/2021
Solution File No.:	00951	Cert. No.: 00032
Sequential File No.:	00951	Lin. Date: 10/25/2021
		Lin. No.: 00031
		Soln. Date: 12/21/2021
		Soln. No.: 00223
		File Date: 12/21/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0756

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:31S		
Control 1 EC	0.098%	10:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33S		
Control 2 EC	0.097%	10:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S		
Control 3 EC	0.098%	10:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DREHER

First Name: BRIAN

MI: M

Signature: 

Badge No.: 166

Date: 12/21/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00869	Calib. Date: 09/01/2020
Certification File No.:	00870	Calib. No.: 00034
Linearity File No.:	00871	Cert. Date: 09/01/2020
Solution File No.:	00887	Cert. No.: 00029
Sequential File No.:	00887	Lin. Date: 09/01/2020
		Lin. No.: 00028
		Soln. Date: 01/04/2021
		Soln. No.: 00201
		File Date: 01/04/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1054

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:04S		
Control 1 EC	0.098%	20:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:06S		
Control 2 EC	0.097%	20:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:07S		
Control 3 EC	0.097%	20:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____



Badge No.: 161

Date: 01/04/2021

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00869	Calib. Date: 09/01/2020
Certification File No.:	00870	Calib. No.: 00034
Linearity File No.:	00871	Cert. Date: 09/01/2020
Solution File No.:	00892	Cert. No.: 00029
Sequential File No.:	00892	Lin. Date: 09/01/2020
		Lin. No.: 00028
		Soln. Date: 02/02/2021
		Soln. No.: 00204
		File Date: 02/02/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1118

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:23S		
Control 1 EC	0.098%	15:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		
Control 2 EC	0.096%	15:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27S		
Control 3 EC	0.097%	15:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	15:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SOFIELD

First Name: RONALD

MI: G

Signature: _____

Sgt. [Signature] #1101

Badge No.: 161

Date: 02/02/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00893	Calib. Date: 02/16/2021
Certification File No.:	00894	Calib. No.: 00035
Linearity File No.:	00895	Cert. Date: 02/16/2021
Solution File No.:	00896	Cert. No.: 00030
Sequential File No.:	00896	Lin. Date: 02/16/2021
		Lin. No.: 00029
		Soln. Date: 02/16/2021
		Soln. No.: 00205
		File Date: 02/16/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1071

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:48S		
Control 1 EC	0.101%	14:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:49S		
Control 2 EC	0.099%	14:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51S		
Control 3 EC	0.097%	14:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDW5P2-227 (DL)

Changed By:

Last Name: LUTZ	First Name: DENNIS	MI: J
Signature: <u>Tpr I Lutz 7045</u>	Badge No.: 7045	Date: 02/16/2021

Calibrating Unit New Standard Solution Report

Equipment
 Location: Alcotest 7110 MKIII-C
 Calibration File No.: 00906
 Certification File No.: 00907
 Linearity File No.: 00908
 Solution File No.: 00911
 Sequential File No.: 00911
 Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 20390

BRIELLE POLICE DEPT.
 Serial No.: ARWM-0073
 Calib. Date: 04/26/2021
 Cert. Date: 04/26/2021
 Lin. Date: 04/26/2021
 Soln. Date: 05/28/2021
 File Date: 05/28/2021
 Model No.: CU-34
 Serial No.: DDXA S3-0052
 Expires: 09/02/2022
 Bottle No.: 1138

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:31D		
Control 1 EC	0.099%	13:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:32D		
Control 2 EC	0.098%	13:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34D		
Control 3 EC	0.098%	13:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: *Stephen Boyd* #156

Badge No.: 156

Date: 05/28/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00906 Calib. Date: 04/26/2021 Calib. No.: 00037
Certification File No.: 00907 Cert. Date: 04/26/2021 Cert. No.: 00031
Linearity File No.: 00908 Lin. Date: 04/26/2021 Lin. No.: 00030
Solution File No.: 00913 Soln. Date: 05/31/2021 Soln. No.: 00213
Sequential File No.: 00913 File Date: 05/31/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0052
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0810

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:30D		
Control 1 EC	0.099%	09:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32D		
Control 2 EC	0.098%	09:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33D		
Control 3 EC	0.098%	09:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: _____



Badge No.: 156

Date: 05/31/2021