

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00937 Calib. Date: 10/25/2021 Calib. No.: 00038
Certification File No.: 00938 Cert. Date: 10/25/2021 Cert. No.: 00032
Linearity File No.: 00939 Lin. Date: 10/25/2021 Lin. No.: 00031
Solution File No.: 00955 Soln. Date: 01/19/2022 Soln. No.: 00224
Sequential File No.: 00955 File Date: 01/19/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0051
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0815

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:42S		
Control 1 EC	0.101%	16:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:43S		
Control 2 EC	0.099%	16:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44S		
Control 3 EC	0.100%	16:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: _____



Badge No.: 156

Date: 01/19/2022