

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00959 Calib. Date: 02/18/2022 Calib. No.: 00039
Certification File No.: 00960 Cert. Date: 02/18/2022 Cert. No.: 00033
Linearity File No.: 00961 Lin. Date: 02/18/2022 Lin. No.: 00032
Solution File No.: 00964 Soln. Date: 03/18/2022 Soln. No.: 00226
Sequential File No.: 00964 File Date: 03/18/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0051
Control Solution %: 0.100% Expires: 12/28/2023
Solution Control Lot: 21480 Bottle No.: 1132

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:25S		
Control 1 EC	0.101%	14:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:26S		
Control 2 EC	0.100%	14:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28S		
Control 3 EC	0.101%	14:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: _____

Lt. Stephen Boyd #156

Badge No.: 156

Date: 03/18/2022