

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00959	Calib. Date: 02/18/2022
Certification File No.:	00960	Calib. No.: 00039
Linearity File No.:	00961	Cert. Date: 02/18/2022
Solution File No.:	00967	Cert. No.: 00033
Sequential File No.:	00967	Lin. Date: 02/18/2022
		Lin. No.: 00032
		Soln. Date: 04/18/2022
		Soln. No.: 00227
		File Date: 04/18/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	20310	Expires: 06/30/2022
		Bottle No.: 1043

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:26D		
Control 1 EC	0.098%	11:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 2 EC	0.096%	11:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 3 EC	0.096%	11:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: _____

Lt Stephen Boyd #156

Badge No.: 156

Date: 04/18/2022