

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0073
Location: BRIELLE POLICE DEPT.
Calibration File No.: 00959 Calib. Date: 02/18/2022 Calib. No.: 00039
Certification File No.: 00960 Cert. Date: 02/18/2022 Cert. No.: 00033
Linearity File No.: 00961 Lin. Date: 02/18/2022 Lin. No.: 00032
Solution File No.: 00972 Soln. Date: 06/14/2022 Soln. No.: 00229
Sequential File No.: 00972 File Date: 06/14/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0051
Control Solution %: 0.100% Expires: 12/28/2023
Solution Control Lot: 21480 Bottle No.: 1082

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:37D		
Control 1 EC	0.100%	10:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:38D		
Control 2 EC	0.099%	10:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39D		
Control 3 EC	0.098%	10:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: Y Williams

First Name: KEVIN

MI: L

Signature: Sgt Williams #163

Badge No.: 163

Date: 06/14/2022