

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0073
Location:	BRIELLE POLICE DEPT.	
Calibration File No.:	00959	Calib. Date: 02/18/2022
Certification File No.:	00960	Calib. No.: 00039
Linearity File No.:	00961	Cert. Date: 02/18/2022
Solution File No.:	00968	Cert. No.: 00033
Sequential File No.:	00968	Lin. Date: 02/18/2022
		Lin. No.: 00032
		Soln. Date: 05/19/2022
		Soln. No.: 00228
		File Date: 05/19/2022

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0051
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 1134

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:13D		
Control 1 EC	0.104%	12:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		
Control 2 EC	0.101%	12:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15D		
Control 3 EC	0.100%	12:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BOYD

First Name: STEPHEN

MI: M

Signature: \_\_\_\_\_

Lt. Stephen Boyd 7756

Badge No.: 156

Date: 05/19/2022