

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Brielle Boro Police Department 601 Union Lane Brielle, NJ 08730		2. Contact: Lt. Stephen Boyd	4. Date: 12/16/2022
		3. Phone Number: (732) 528-5056	5. County: Monmouth
6. Alcotest Instrument Serial Number: N/A	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
I responded to the Brielle Police Department to complete the semi-annual recalibration of the Alcotest 9510.

10. Comments:
On December 16, 2022 I responded to the Brielle Borough Police Department to complete the semi-annual recalibration of the Alcotest 9510. I attempted to complete the certification procedure and discovered the gateway (serial number DRRF-0423) was not communicating with the Alcotest 9510. I checked all wires were connected properly and power cycled the gateway and the Alcotest. These attempts were unable to restore the communication between gateway and Alcotest. Due to the lack of communication between the gateway and Alcotest, I could not complete the certification procedure. The Alcotest was placed out of service and the gateway (serial number DRRF-0423) was returned to Draeger Inc. for evaluation.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p>Sgt. Kevin W. Alcott #6704 Name & Badge Number (Print)</p>	<p><i>Sgt. K W. A #6704</i> Signature</p>	<p>12/16/2022 Date</p>
---	---	----------------------------