

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Brielle Boro Police Department 601 Union Lane Brielle, NJ 08730	2. Contact: Lt. Stephen Boyd 3. Phone Number: (732) 528-5056	4. Date: 03/29/2023 5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">N/A</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
The gateway returned from Draeger Inc.

10. Comments:
The gateway (serial number DRRF-0423) was placed back in service.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr II. Nicholas E. Mimikos #7413
Name & Badge Number (Print)

Tpr II. N.E. #7413
Signature

03/29/2023
Date